

SKC Noisechek Dosimeter

Visual Inspection	Yes	No
Physical Damage to Instrument / Screen?	<input type="checkbox"/>	<input type="checkbox"/>
Physical Damage to Docks / Acoustichek Calibrator?	<input type="checkbox"/>	<input type="checkbox"/>

Calibration & Functional Inspection	Yes	No
Pre-Calibrated?	<input type="checkbox"/>	<input type="checkbox"/>
Varying Sound Level Readings on Display?	<input type="checkbox"/>	<input type="checkbox"/>
ALL Buttons Function Properly?	<input type="checkbox"/>	<input type="checkbox"/>

All Dosimeter Calibrations Valid through Rental End Date?	Y	N
	<input type="checkbox"/>	<input type="checkbox"/>

Setup & Data	Y	N
Standard Set-up with DataTrac Software OSHA HC, OSHA PEL, ACGIH? See back of checklist for full configuration.	<input type="checkbox"/>	<input type="checkbox"/>
Date/Time Set and Data Cleared	<input type="checkbox"/>	<input type="checkbox"/>
Auto Lock / Secure lock DISABLED?	<input type="checkbox"/>	<input type="checkbox"/>
Firmware up to date?	<input type="checkbox"/>	<input type="checkbox"/>
Battery charged? (75% Minimum)	<input type="checkbox"/>	<input type="checkbox"/>

What's In the Box?
An Inspection Guide to the RAECO Rents Instrumentation You Requested.

The contents of your rental kit will contain all the components shown on this sheet. When you unpack the instrument, please try turning it on. If nothing happens, or a kit component is missing, please call (866)-736-8347 immediately so we can correct

Not Shown: Spare AA Batteries

Included Parts & Accessories	Yes	No
SKC Noisechek Dosimeter(s) QTY: _____	<input type="checkbox"/>	<input type="checkbox"/>
(Single Bay / Five Bay) Dock QTY: 5-Bay _____ QTY: Single Bay _____	<input type="checkbox"/>	<input type="checkbox"/>
A/C Power Cable QTY: _____	<input type="checkbox"/>	<input type="checkbox"/>
SKC Acoustichek Calibrator	<input type="checkbox"/>	<input type="checkbox"/>
Micro USB Cable QTY: _____	<input type="checkbox"/>	<input type="checkbox"/>
Spare Windscreen QTY: _____	<input type="checkbox"/>	<input type="checkbox"/>
Spare Clip with Pin: QTY _____	<input type="checkbox"/>	<input type="checkbox"/>
Spare (2) AA Batteries for Acoustichek Calibrator	<input type="checkbox"/>	<input type="checkbox"/>
DataTrac dB Software, Dosimeter and Calibrator User Manuals / Quick Start Guides – May be on Flash Drive Only	<input type="checkbox"/>	<input type="checkbox"/>



***** Please note *****

PRESS STRAIGHT DOWN ON BUTTONS. DO NOT PRESS AT AN ANGLE. They are not designed to be pressed at an angle and may be damaged if pressed incorrectly often.

Additional Parts or Customer Setup Requests	Yes	No
CUSTOM CUSTOMER SETUP? See notes on contract for setup instructions		
Octave Band Enabled? 1/1 or 1/3	<input type="checkbox"/>	<input type="checkbox"/>
Sound Recording Activated? Threshold for Recording = _____ dB	<input type="checkbox"/>	<input type="checkbox"/>

Instrument Technician: _____

Interested in purchasing this instrument? Call us for a quote! (866)736-8347

Date: _____ / _____

SKC Noisechek Dosimeter

Standard RAECO Rents Dosimeter Configuration

Setup
Schedule
History

Select up to nine readings to show on the device during a run

<input checked="" type="checkbox"/> SPL	<input type="checkbox"/> Lmin	<input type="checkbox"/> C-A
<input checked="" type="checkbox"/> TWA	<input type="checkbox"/> Exposure	<input type="checkbox"/> LEP,d
<input checked="" type="checkbox"/> Dose	<input checked="" type="checkbox"/> pTWA	<input type="checkbox"/> LEX,8h
<input checked="" type="checkbox"/> Lavg	<input checked="" type="checkbox"/> pDose	<input type="checkbox"/> Exposure Pts
<input checked="" type="checkbox"/> Peak	<input type="checkbox"/> SEL	<input type="checkbox"/> Exposure Pts/Hr
<input checked="" type="checkbox"/> Lmax	<input checked="" type="checkbox"/> Upper Limit	<input type="checkbox"/> CUL

Select up to seven readings to show on the device in History

<input checked="" type="checkbox"/> TWA	<input type="checkbox"/> Exposure	<input type="checkbox"/> LEP,d
<input checked="" type="checkbox"/> Dose	<input checked="" type="checkbox"/> pTWA	<input type="checkbox"/> LEX,8h
<input checked="" type="checkbox"/> Lavg	<input checked="" type="checkbox"/> pDose	<input type="checkbox"/> Exposure Pts
<input checked="" type="checkbox"/> Peak	<input type="checkbox"/> SEL	<input type="checkbox"/> Exposure Pts/Hr
<input checked="" type="checkbox"/> Lmax	<input type="checkbox"/> Upper Limit	<input type="checkbox"/> CUL
<input type="checkbox"/> Lmin	<input type="checkbox"/> C-A	

Set other options on the device

<input type="checkbox"/> Display octave bands	<input checked="" type="checkbox"/> Log data	<input type="checkbox"/> Secure Lock
<input type="checkbox"/> Log octave band data	<input type="text" value="1 sec"/> <input type="text" value="60 sec"/>	Require PIN to connect to mobile app
<input type="text" value="1 octave"/> <input type="text" value="1/3 octave"/>	Peak Weighting	PIN <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>
<input type="checkbox"/> Disable Voice Notes	<input type="text" value="C"/> <input type="text" value="Z"/>	<input type="checkbox"/> Auto Lock Require PIN to stop or pause

Enable and define up to four separate virtual dosimeters for each run

OSHA - HC	×	OSHA - PEL	×	ACGIH	×	
Response	Slow	Response	Slow	Response	Slow	
Exchange Rate	5 dB	Exchange Rate	5 dB	Exchange Rate	3 dB	
Threshold	80 dB	Threshold	90 dB	Threshold	80 dB	
Criterion Level	90 dB	Criterion Level	90 dB	Criterion Level	85 dB	
Weighting	A	Weighting	A	Weighting	A	
Upper Limit	115 dB	Upper Limit	115 dB	Upper Limit	115 dB	

<input type="checkbox"/> Auto-record at <input type="text" value="0"/> dB	CUL Threshold <input type="text" value="117"/> dB	pTWA/pDose Time <input type="text" value="8"/> hr
<input type="checkbox"/> Alert at <input type="text" value="0"/> % Dose	CUL Interval <input type="text" value="30"/> sec	

Instrument Technician: _____

Interested in purchasing this instrument? Call us for a quote! (866)736-8347

Date: _____ / _____